

VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

OCT 26 2009

DataMaster cdm S/N: 130584

Supervisor/Operator Performing the Verification Procedure:

Name Supt. James R. R. Rea

ID# 3161

Date 10-05-09

A Agency Kotzebue Police Department

Phone # 907-442-3352

Instrument Location Kotzebue Regional Jail 258d rm#107 Fourth Avenue Kotzebue, AK 99752

Alco S/N 78890

Target Value .082

High Pressure 1400

Alco Test Values:

.082
1st Alco

.083
2nd Alco

B

Signature _____

Supt. James R. R. Rea
(OVER)

CRB
10/4/09

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 21st day of Dec, 2009.

Carolyn M. Noland
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130584

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

Breath Test Record

Serial Number 130584 Model DataMaster
Location DTZ1 Kotzebue Police Dept.

Subject Information

Last Name Verification Of Cal O.L. # 10052009
First Name Verification Of Cal

Operator Information

Operator Name Rea/James/Rr Case 10052009
Operator # 3161 Department DTZ1

Test Results

Alco Target 082 Alco Serial 78890
Nominal 0.082 Barometer 30.20 Test Type V
Dry Gas 1 0.082 Time: 12:58:38 Test Date 10/05/2009
Subject Alcohol 0.000 Time: 12:59:46 Test Time 12:56:53
Dry Gas 2 0.083 Time: 13:01:00 Report Y

.082
ALCO 2:56
BLANK 2:58
INTER 2:58
ALCO 2:58
BLANK 2:59
SUBJE 2:59
BLANK 2:59
ALCO 2:59
BLANK TEST 13:01

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130584

Diagnostic Test Record

Serial Number 130584 Model DataMaster
Location DTZ1 Kotzebue Police Dept.

Diagnostic Test Results

Test Date 10/05/2009 Computer Okay
Test Time 14:00:49 Program Okay
Software Date 02/20/2001
Sample Chamber (C) 49
Breath Tube (C) 41
Barometer (in) 30.22
Flow Detector Okay
High Speed Pump Okay
Detector Okay
Filter Okay
Quartz Okay
Calibration Okay

DETECT
FILTER
QUARTZ
CALIBR
!###
HIJKLMNOPQRSTUWXYZ\]^_`abcde{ghijklmno
pqrstuvwxyz{|}+~